

Attestation of Conformity



Korea

Manufacturer **i-SENS, Inc.**
27-36, Gwangun-ro, Nowon-gu
Seoul 139-845, Republic of Korea

Item Identification **alphacheck professional**
Blood Glucose Monitoring System
(Meter Model: GM505RA)

Above mentioned product manufactured by i-SENS, Inc.,
meets the Minimum System Accuracy Performance Criteria,
defined in the clause 6.3.3 of ISO 15197:2013.

This attestation is released based on the Technical File Assessment Result as they
submitted a Change Notification regarding above mentioned product. The detailed
assessment results are listed in

Test Report No. **74939333** (Change Notification ID: IS-TED-016)

Issue Place:

Medical Health Services
of TÜV SÜD Korea Ltd.
Subsidiary of
TÜV SÜD Product Service GmbH

A handwritten signature in blue ink, appearing to read 'T. Ushiki', followed by the date '2014-6-25'.

Dr. Takeo Ushiki
Manager of
Medical Health Services